

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、
当該導入に関して資料等の提供を招請します。

令和7年6月6日

契約担当役

国立大学法人千葉大学事務局長 丸山 浩

◎調達機関番号 415 ◎所在地番号 12

○第1号

1 調達内容

(1) 品目分類番号 22、31

(2) 導入計画物品及び数量

ナースコール管理システム 一式

(3) 調達方法 購入等

(4) 導入予定時期 令和8年7月

(5) 調達に必要とされる基本的な要求要件

①システム構成

- ・ナースコールシステム
- ・ベッドサイド、トイレ、浴室用子機
- ・見守りカメラ
- ・トイレ離座センサー
- ・徘徊センサー

②本院の診療業務、管理運営に影響を与えない円滑な動作品質を、24時間365日提供できること。

③本院の診療業務、管理運営を総合的に支援できる機能を実現すること。

④電子カルテと連携し、リアルタイムに患者情報等を反映できること。

⑤診療業務に支障なく、高速に配信・表示・処理ができること。

⑥システムの障害を未然に検知・防止するとともに、障害発生時には即時対応できるサポート体制を有すること。

⑦一部の機器が停止しても、システム全体の機能が消失することがないシステム構成であること。

⑧省エネルギー、省スペースを十分に考慮したシステム構成であること。

⑨管理運用が容易で、適宜設定変更を行えること。

2 資料及びコメントの提供方法 上記1(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

(1) 資料等の提供期限 令和7年7月25日17時00分（郵送の場合は必着のこと。）

(2) 提供先 〒260—8677 千葉市中央区亥鼻1—8—1 千葉大学医学部附属病院管理課調達グループ 佐々木 結花 電話 043—226—2236

3 説明書の交付 本公表に基づき応募する供給者に対して導入説明書を交付する。

(1) 交付期間 令和7年6月6日から令和7年7月25日まで。

(2) 交付場所 上記2(2)に同じ。

4 説明会の開催 本公表に基づく導入説明会を開催する。

(1) 開催日時 令和7年6月11日11時00分

(2) 開催場所 千葉大学医学部附属病院外来診療A棟3階セミナー室1

5 その他 この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6 Summary

(1) Classification of the products to be procured : 22, 31

(2) Nature and quantity of the products to be purchased : Nurse Call Management System 1 set

(3) Type of the procurement : Purchase

(4) Basic requirements of the procurement :

① System configuration

- Nurse call system
- Handset for bedside, toilet, and bathroom
- Monitoring camera
- Toilet leaving sensor
- Wandering sensor

② To be able to provide smooth operational quality 24 hours a day, 365 days a year, without affecting the medical services, management and administration of this hospital.

③ Realize a function that can comprehensively support the medical care work, management and operation of this hospital.

④ Electronic health records can be integrated to reflect patient information in real-time.

⑤ Ability to deliver, display, and process at high speed without interfering with medical services.

⑥ To detect and prevent system failures, and to have a support system that can respond immediately in the event of a failure.

⑦ The system configuration shall be such that even if a part of the equipment stops, the functions of the entire

system will not be lost.

⑧ The system configuration should fully consider energy saving and space saving.

⑨ Management and operation should be easy, and settings should be changed as appropriate.

(5) Time limit for the submission of the requested material : 17:00 25 July, 2025

(6) Contact point for the notice : Yuka Sasaki, Procurement Group, Management Division, Chiba University Hospital, 1—8—1 Inohana Chuo-ku Chiba-shi 260—8677 Japan, TEL 043—226—2236